spectra® laboratories
525 Sycamore Dr
Milpitas, California 95035
800-433-ESRD (3773)

Patient Name
WATER,TDC/EMILE MOHAMMED

Sex:

RENAL DYNAMICS LLC (W) <75663>

ATTN:CHRIS DAVENPORT
3020 KANANWOOD CT, STE 1000
OVIEDO,FL 32765

Physician:
Account I.D.:
Specimen No.: 7255624

Date of Specimen: 10/14/2009
Date Received: 10/19/2009
Date Reported: 10/24/2009
Time Reported: 01:15AM

REQUISITION NUMBER: Y775744
SAMPLE: WATER
SOURCE: RO
COLLECTED BY: JAMES MCWHORTER 407 333 0028
COMMENTS: POST RO
RO
THE DIALYSIS CLINIC BARBADOS
DR. EMILE MOHAMMED

TEST	**RESULT**	**UNITS**	**REFERENCE**
WATER ANALYSIS			
SODIUM,WATER	2.579	mg/L	0-70
POTASSIUM,WATER	<1.000	mg/L	0-8
ALUMINUM,WATER	<0.008	mg/L	0-0.01
CALCIUM,WATER	<0.050	mg/L	0-2
COPPER,WATER	<0.005	mg/L	0-0.1
MAGNESIUM,WATER	<0.050	mg/L	0-4
SELENIUM,WATER	<0.005	mg/L	0-0.09
ZINC,WATER	<0.005	mg/L	0-0.1
CHROMIUM,WATER	<0.005	mg/L	0-0.014
LEAD,WATER	<0.002	mg/L	0-0.005
ARSENIC,WATER	<0.002	mg/L	0-0.005
MERCURY,WATER	<0.0002	mg/L	0-0.0002
CADMIUM,WATER	<0.0010	mg/L	0-0.001
BERYLLIUM,WATER	<0.0004	mg/L	0-0.0004
ANTIMONY,WATER	<0.006	mg/L	0-0.006
THALLIUM,WATER	<0.002	mg/L	0-0.002
SILVER,WATER	<0.003	mg/L	0-0.005
BARIUM,WATER	<0.001	mg/L	0-0.1
FLUORIDE,WATER	<0.10	mg/L	0-0.20
NITRATE (N),WATER	0.2	mg/L	0-2.0
SULFATE,WATER	<1.0	mg/L	0-100
PH	6.7	UNITS	
CONDUCTIVITY	23.50	uS/cm	
RESISTIVITY	0.043	MEG OHMS	
WATER COMMENT:			

WATER TESTED TO AAMI STANDARDS

FINAL REPORT ---- DO NOT DESTROY

WATER,TDC/EMILE MOHAMMED

LABORATORY DIRECTOR: Claude O. Burdick, M.D.